



湖北省计量测试技术研究院

Hubei Institute of Measurement and Testing Technology

校准证书

Calibration Certificate

证书编号: 2026RG046600015
Certificate No.

委托方
Client

上汽通用五菱汽车股份有限公司青岛分公司

委托方地址
Address

青岛市黄岛区江山中路1号

器具名称
Name of instrument

全自动运动粘度测定仪

制造厂商
Manufacturer

武汉特高压电力科技有限公司

型号/规格
Type/Specification

UHV-675

器具编号
Serial No.

2601017

湖北省计量测试
证书骑

批准人
Approved by

张玉婷

张玉婷

核验员
Checked by

李建欢

李建欢

校准员
Calibrated by

樊宁君

樊宁君



样品接收日期(1)

Date of Application

2026 年 01 月 06 日

校准日期

Date of Calibration

2026 年 01 月 12 日

签发日期

Date of Issue

2026 年 01 月 13 日



地址: 湖北省武汉市东湖新技术开发区茅店山中路二号(总部)

Add: No.2, Maodianshanzhong Road, East Lake High-tech Development Zone, Wuhan, Hubei

网址(Web site): <http://www.himtt.net>

邮编(Post Code): 430223

电话(Tel): 027-81925136

传真(Fax): 027-81925137

第 1 页共 3 页

Page of total pages

Z260100181

Z260100181-1-001



扫描全能王 创建

● 本院是政府计量行政管理部门依法设立的国家法定计量检定机构

This laboratory is a legal metrological verification institution established by the government metrological administrative department according to law.

● 本次校准的技术依据 (代号、名称)

Reference documents for the Calibration (Code, Name)

参照 JJF 2019-2022 《液体恒温试验设备温度性能测试规范》

● 本次校准所使用的主要计量标准器具

Main standards of measurement used in the Calibration

名称 Name	型号/编号 Type/Serial No.	测量范围 Measuring Range	不确定度/准确度等级/ 最大允许误差 Uncertainty/Accuracy Class/ Maximum Permissible Error	证书编号/ 溯源机构 Certificate No / Traceability Agency	有效期限 Due date
温湿度巡检仪	CF32 AB-2204044	温度 (-80~300) ℃湿度 (10-90) % RH	铂电阻MPE:±0.10℃; 热电偶MP E:±0.4%; 湿度MPE:±1.0%RH.	2025RG046600075 湖北省计量测试 技术研究院	2026-02-13

● 校准地点及环境条件

Environmental condition and Place for the Calibration

地 点: 本院水果湖基地B203

Site

温 度: 19 ℃

Temperature

相对湿度: 48 %

R.H.

其 它: /

Others

气 压: / kPa

Pressure

原始记录编号: 2026RG046600015

Record No.

● 备注 /

Note

1、本院所出具的量值 (校准结果), 均可溯源至国家计量基准和 (或) 国际单位制 (SI)。

All data issued by this laboratory are traceable to national primary standards an international system of units(SI).

2、校准结果, 仅对受校样品的本次校准有效。校准证书封面未加盖校准专用章无效。

It's Effect That Results of This Report Relate Only To The Sample(s) Calibrated. It's Invalid That The Certificate Cannot Be Stamped.

3、未经本院书面批准, 不得复制 (全文复制除外)。

Without the written approval of the court, it is not allowed to copy (except full-text copy)

4、“委托方”、“委托方联络信息”由委托方提供, “制造厂”、“型号规格”、“器具编号”为仪器上标注。

The information Client and Contact Information are provided by client, and the Manufacturer, Type /Specification, Serial No. are marked on the items.

校准证书专用续页

Continued page of calibration certificate

第 2 页共 3 页

Page of total pages



扫描全能王 创建

校准结果
Results of Calibration

设定温度 (℃)	50	扩展不确定度
温度上偏差 (℃)	-0.5	$U=0.5^{\circ}\text{C}, k=2$
温度下偏差 (℃)	-0.7	$U=0.5^{\circ}\text{C}, k=2$

以下空白

